

# T3<sup>®</sup> PRO, T3<sup>®</sup>, and Osseotite<sup>®</sup> Tapered Implant Surgical Kit Reusable Drill Usage Chart



Use a dry erase marker to track the total number of osteotomy preparations that are performed with each drill. It is recommended that reusable drills be replaced after 15 uses.\* Single use drills should be disposed of after each procedure.



	ACTPSD	ACT2010	ACT2015	ACT2020	ICD100	QSD3285	QSD3210	QSD3211	QSD3213	QSD3215	QSD485	QSD410	QSD411	QSD413	QSD415	QSD585	QSD510	QSD511	QSD513	QSD515	QSD685	QSD610	QSD611	QSD613	QSD615
	ACT <sup>®</sup> Pointed Starter Drill	ACT Twist Drill 2 mm x 10 mm	ACT Twist Drill 2 mm x 15 mm	ACT Twist Drill 2 mm x 20 mm	Countersink Drill 4 mm	Quad Shaping Drill 8.5 mm	Quad Shaping Drill 10 mm	Quad Shaping Drill 11.5 mm	Quad Shaping Drill 13 mm	Quad Shaping Drill 15 mm	Quad Shaping Drill 8.5 mm	Quad Shaping Drill 10 mm	Quad Shaping Drill 11.5 mm	Quad Shaping Drill 13 mm	Quad Shaping Drill 15 mm	Quad Shaping Drill 8.5 mm	Quad Shaping Drill 10 mm	Quad Shaping Drill 11.5 mm	Quad Shaping Drill 13 mm	Quad Shaping Drill 15 mm	Quad Shaping Drill 8.5 mm	Quad Shaping Drill 10 mm	Quad Shaping Drill 11.5 mm	Quad Shaping Drill 13 mm	Quad Shaping Drill 15 mm
						3.25 mmD					4 mmD					5 mmD					6 mmD				
1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2
3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3	3
4	4	4	4	4	4	4	4	4	4	4	4	4	4	4	4	4	4	4	4	4	4	4	4	4	4
5	5	5	5	5	5	5	5	5	5	5	5	5	5	5	5	5	5	5	5	5	5	5	5	5	5
6	6	6	6	6	6	6	6	6	6	6	6	6	6	6	6	6	6	6	6	6	6	6	6	6	6
7	7	7	7	7	7	7	7	7	7	7	7	7	7	7	7	7	7	7	7	7	7	7	7	7	7
8	8	8	8	8	8	8	8	8	8	8	8	8	8	8	8	8	8	8	8	8	8	8	8	8	8
9	9	9	9	9	9	9	9	9	9	9	9	9	9	9	9	9	9	9	9	9	9	9	9	9	9
10	10	10	10	10	10	10	10	10	10	10	10	10	10	10	10	10	10	10	10	10	10	10	10	10	10
11	11	11	11	11	11	11	11	11	11	11	11	11	11	11	11	11	11	11	11	11	11	11	11	11	11
12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12
13	13	13	13	13	13	13	13	13	13	13	13	13	13	13	13	13	13	13	13	13	13	13	13	13	13
14	14	14	14	14	14	14	14	14	14	14	14	14	14	14	14	14	14	14	14	14	14	14	14	14	14
15	15	15	15	15	15	15	15	15	15	15	15	15	15	15	15	15	15	15	15	15	15	15	15	15	15

Time to Order New Drills

\* This chart is intended to be used as a guide only. Improper drill use may lead to premature wear and the need to replace the drill prior to 15 uses.

**Note:** Surgical instruments are susceptible to damage for a variety of reasons including but not limited to: prolonged use, misuse, rough or improper handling, inadequate cleaning, etc. Care must be taken to avoid compromising the intended performance of the instrument. Please refer to P-IFSCSS for complete instructions on the sterilization and care of stainless steel.

For more information, visit [ZimVie.com](http://ZimVie.com)

Unless otherwise indicated, as referenced herein, all trademarks are the property of ZimVie; and all products are manufactured by one or more of the dental subsidiaries of ZimVie Inc. (Biomet 3i, LLC, Zimmer Dental Inc., etc.) and marketed and distributed by ZimVie Dental and its authorized marketing partners. For additional product information, please refer to the individual product labeling or instructions for use. Product clearance and availability may be limited to certain countries/regions. This material is intended for clinicians and Dental Laboratory use only and does not comprise medical advice or recommendations. Distribution to any other recipient is prohibited. This material may not be copied or reprinted without the express written consent of ZimVie Inc. ZV1201 REV A 05/23 ©2023 ZimVie Inc. All rights reserved.



# Know How to Identify Excessive Wear

## Inspection for Wear and Damage

Before each use, visually inspect all devices and trays/kits for completeness, damage and/or excessive wear (e.g. corrosion or rust build-up on the instrument surface), structural wear or damage, grommet presence or damage, partial or complete fracture. If damage or wear is noted that may compromise the function of the instrument or tray, **DO NOT USE**. Contact your ZimVie representative for replacement.

Refer to Instructions for Use (IFU) for Biomet 3i Kits and Instruments (P-ZBDINSTRP) available at [labeling.zimvie.com](http://labeling.zimvie.com) for complete Cleaning and Sterilization information.

